

New England Biolabs Certificate of Analysis

Product Name: Bsal-HF@v2
Catalog Number: R3733S
Concentration: 20,000 U/ml
Unit Definition: One unit is defined as the amount of enzyme required to digest 1 µg of pXba DNA in 1 hour at 37°C in a total reaction volume of 50 µl.
Packaging Lot Number: 10110101
Expiration Date: 03/2023
Storage Temperature: -20°C
Storage Conditions: 10 mM Tris-HCl, 200 mM NaCl, 1 mM DTT, 0.1 mM EDTA, 200 µg/ml BSA, 50 % Glycerol, (pH 7.4 @ 25°C)
Specification Version: PS-R3733S/L v1.0

Bsal-HF@v2 Component List			
NEB Part Number	Component Description	Lot Number	Individual QC Result
R3733SVIAL	Bsal-HF@v2	10100710	Pass
B7024AVIAL	Gel Loading Dye, Purple (6X)	10093123	Pass
B6004SVIAL	rCutSmart™ Buffer	10108412	Pass

Assay Name/Specification	Lot # 10110101
Exonuclease Activity (Radioactivity Release) A 50 µl reaction in CutSmart® Buffer containing 1 µg of a mixture of single and double-stranded [³ H] E. coli DNA and a minimum of 100 units of Bsal-HF@v2 incubated for 4 hours at 37°C releases <0.1% of the total radioactivity.	Pass
Endonuclease Activity (Nicking) A 50 µl reaction in CutSmart® Buffer containing 1 µg of supercoiled PhiX174 DNA and a minimum of 20 units of Bsal-HF@v2 incubated for 4 hours at 37°C results in <20% conversion to the nicked form as determined by agarose gel electrophoresis.	Pass
Non-Specific DNase Activity (16 Hour) A 50 µl reaction in CutSmart® Buffer containing 1 µg of pXba DNA and a minimum of 60 units of Bsal-HF@v2 incubated for 16 hours at 37°C results in a DNA pattern free of detectable nuclease degradation as determined by agarose gel electrophoresis.	Pass
Ligation and Recutting (Terminal Integrity) After a 20-fold over-digestion of pXba DNA with Bsal-HF@v2, >95% of the DNA fragments can be ligated with T4 DNA ligase in 16 hours at 16°C. Of these ligated	Pass

Assay Name/Specification	Lot # 10110101
<p>fragments, >95% can be recut with Bsal-HF@v2.</p> <p>Protein Purity Assay (SDS-PAGE) Bsal-HF@v2 is \geq 95% pure as determined by SDS-PAGE analysis using Coomassie Blue detection.</p>	<p>Pass</p>

This product has been tested and shown to be in compliance with all specifications.

One or more products referenced in this document may be covered by a 3rd-party trademark. Please visit www.neb.com/trademarks for additional information.



Penghua Zhang
Production Scientist
14 Jun 2021



Michael Tonello
Packaging Quality Control Inspector
14 Jun 2021